UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.	
10/556,918	11/14/2005	Carl J. Brunnett	PHUS030122ÜS	1574	
38107 7590 12/28/2007 PHILIPS INTELLECTUAL PROPERTY & STANDARDS 595 MINER ROAD			EXAM	EXAMINER	
			· MIDKIFF, A	MIDKIFF, ANASTASIA	
CLEVELAND	OH 44143	ART UNIT PAPER NUMBER		PAPER NUMBER	
			2882		
			. MAIL DATE	DELIVERY MODE	
•			12/28/2007	PAPER	

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

L		Application No.	Applicant(s)			
Office Action Summary		10/556,918	BRUNNETT, CARL J.			
		Examiner	Art Unit			
		Anastasia Midkiff	2882			
	The MAILING DATE of this communication app		orrespondence address			
Period fo	• •					
WHIC - Exter after - If NO - Failu Any	ORTENED STATUTORY PERIOD FOR REPLY CHEVER IS LONGER, FROM THE MAILING DANS IN THE MAIL	ATE OF THIS COMMUNICATION 36(a). In no event, however, may a reply be timused and will expire SIX (6) MONTHS from a cause the application to become ABANDONE	N. nely filed the mailing date of this communication. D (35 U.S.C. § 133).			
Status						
1)⊠	Responsive to communication(s) filed on 30 No	<u>ovember 2007</u> .				
,—	This action is FINAL : 2b)⊠ This action is non-final.					
3)	•					
	closed in accordance with the practice under E	х рапе Quayle, 1935 С.D. 11, 48	33 O.G. 213.			
Dispositi	ion of Claims					
4)⊠	4) Claim(s) <u>1-20</u> is/are pending in the application.					
	4a) Of the above claim(s) is/are withdrawn from consideration.					
·	Claim(s) is/are allowed.					
· <u> </u>	Claim(s) <u>1-6</u> is/are rejected.					
-	Claim(s) <u>7-20</u> is/are objected to. Claim(s) are subject to restriction and/o	r election requirement				
٥/١	are subject to restriction and/o	·				
Applicati	ion Papers					
• —	The specification is objected to by the Examine					
10)	The drawing(s) filed on is/are: a) acce					
	Applicant may not request that any objection to the					
11)	Replacement drawing sheet(s) including the correct The oath or declaration is objected to by the Ex					
יייי	The ball of declaration is objected to by the Ex		Action of formal 10-132.			
Priority (under 35 U.S.C. § 119					
	Acknowledgment is made of a claim for foreign	priority under 35 U.S.C. § 119(a))-(d) or (f).			
a)	☐ All b)☐ Some * c)☐ None of:					
	1. Certified copies of the priority document		ion No			
	2. Certified copies of the priority documents3. Copies of the certified copies of the priority					
	 Copies of the certified copies of the prior application from the International Bureau 		su in this National Stage			
* 5	See the attached detailed Office action for a list		ed.			
Attachmen		4) Interview Summary	(PTO.413)			
2) Notic	ce of References Cited (PTO-892) ce of Draftsperson's Patent Drawing Review (PTO-948)	Paper No(s)/Mail D	ate			
	mation Disclosure Statement(s) (PTO/SB/08) er No(s)/Mail Date	5) Notice of Informal F 6) Other:	'atent Application			

10/556,918 Art Unit: 2882

DETAILED ACTION

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claims 1-6 are rejected under 35 U.S.C. 103(a) as being unpatentable over U.S. Patent to Trotel (US 5,022,060) in view of U.S. Patent to Yu (US 6,094,473), and further in view of U.S. Patent to Brunnett (US 4,052,620).

With respect to Claim 1, Trotel teaches a CT scanner (Column 6, Lines 37-44), comprising:

- a support (6) for rotating a radiation source (2) around an examination region (Column 6, Lines 16-26); and
- a radiation detector (4) comprising an ion chamber (Column 2, Lines 20-34) for generating an analog signal that varies with an intensity of radiation traversing the examination region (Column 2, Lines 25-36).

Trotel does not specifically teach a means for converting an analog data signal to a digital data signal including aperiodic pulses varying in frequency with the intensity of the radiation traversing the examination region as the radiation source rotates about the examination region, a means for producing a time signal indicative of data intervals, or a means for determining average radiation intensity in each data interval by counting the

10/556,918 Art Unit: 2882

pulses of the digital data signal starting with a digital data signal pulse occurring in a preceding data interval and continuing to a digital data signal pulse occurring in a succeeding data interval.

Yu teaches an x-ray scanner (1) with an ion chamber detector (16), and the method for its use, wherein there is provided:

- a plurality of x-ray detector sensors (30a-c) for generating an analog data signal that varies with an intensity of radiation traversing the examination region;
- a digital frequency modulated output circuit (40) with a converter circuit
 (44) for converting the analog data signal to a digital data signal including aperiodic pulses varying in frequency with the intensity of radiation
 traversing the examination region (Abstract, Lines 1-6);
- a means (40, 44, 46, 60, 70) for producing a time signal indicative of data intervals (Column 6, Lines 1-8);
- a means (60, 70, 72) for determining radiation intensity in a data interval including a processor circuit (72) by counting the pulses of the digital data signal in counter circuits (60, 70), starting with a digital data signal pulse occurring in a preceding data interval and continuing to a digital data signal pulse occurring in a succeeding data interval, said intervals stored in the counter register (Column 6, Lines 15-64)

to provide improved signal-to-noise ratio (Column 3, Lines 64-67).

It would have been obvious to one of ordinary skill in the art at the time of the rejection to use the frequency modulated system of Yu in the apparatus and method of Trotel, to improve signal to noise ratio in signals produced during the scanning of Trotel, as demonstrated by Yu (Abstract and Column 3, Lines 64-67), thereby improving image quality.

Trotel and Yu do not teach averaging radiation intensity in each data interval.

Brunnett teaches an x-ray CT scanner (Column 8, Lines 17-19) in which the reconstruction processor (96) includes a comparator (124) that produces average radiation intensity (Column 7, Lines 9-22) from pulse counts collected by a pulse counting circuit (112) for time periods coinciding with primary time period data intervals (Column 6, Lines 65-68, and Column 7, Lines 1-22) to prevent statistical error in data collection due to any variations in extent of the data intervals (Column 5, Lines 43-52, and Column 6, Lines 18-24).

It would have been obvious to one of ordinary skill in the art at the time of the invention to further modify the apparatus of Trotel to include the data interval radiation intensity averaging suggested by Brunnett, to improve statistical accuracy of data collection, as suggested by Brunnett (Column 6, Lines 17-28, and Column 7, Lines 55-57) for improved imaging.

With respect to Claim 6, Trotel teaches a method of measuring an intensity of detected radiation in a CT scanner (Column 6, Lines 37-44), the method comprising:

rotating a radiation source (2) around an examination region (Column 6,
 Lines 16-26); and,

10/556,918 Art Unit: 2882

> generating an analog signal that varies with an intensity of radiation traversing the examination region (Column 2, Lines 25-36).

Trotel does not specifically teach converting an analog data signal to a digital data signal including aperiodic pulses varying in frequency with the intensity of the radiation traversing the examination region, producing a time signal indicative of data intervals, or determining average radiation intensity in each data interval by counting the pulses of the digital data signal starting with a digital data signal pulse occurring in a preceding data interval and continuing to a digital data signal pulse occurring in a succeeding data interval.

Yu teaches a method of detecting radiation with an x-ray scanner (1) with an ion chamber detector (16), wherein the method includes the steps of:

- generating an analog data signal that varies with an intensity of radiation traversing the examination region with a plurality of detector sensors (30ac);
- converting the analog data signal to a digital data signal including
 aperiodic pulses varying in frequency with the intensity of radiation
 traversing the examination region (Abstract, Lines 1-6) by using a digital
 frequency modulated output circuit (40) with a converter circuit (44);
- producing a time signal indicative of data intervals (Column 6, Lines 1-8);
 and.
- determining radiation intensity in a data interval including a processor
 circuit (72) by counting the pulses of the digital data signal in counter

10/556,918 Art Unit: 2882

circuits (60, 70), starting with a digital data signal pulse occurring in a preceding data interval and continuing to a digital data signal pulse occurring in a succeeding data interval, said intervals stored in the counter register (Column 6, Lines 15-64)

to provide improved signal-to-noise ratio (Column 3, Lines 64-67).

It would have been obvious to one of ordinary skill in the art at the time of the rejection to use the frequency modulated system of Yu in the apparatus and method of Trotel, to improve signal to noise ratio in signals produced during the scanning of Trotel, as demonstrated by Yu (Abstract and Column 3, Lines 64-67), thereby improving image quality.

Trotel and Yu do not teach averaging radiation intensity in each data interval.

Brunnett teaches a method of detecting radiation in an x-ray CT scanner (Column 8, Lines 17-19) in which the reconstruction processor (96) includes a comparator (124) that produces average radiation intensity (Column 7, Lines 9-22) from pulse counts collected by a pulse counting circuit (112) for time periods coinciding with primary time period data intervals (Column 6, Lines 65-68, and Column 7, Lines 1-22) to prevent statistical error in data collection due to any variations in extent of the data intervals (Column 5, Lines 43-52, and Column 6, Lines 18-24).

It would have been obvious to one of ordinary skill in the art at the time of the invention to further modify the apparatus of Trotel to include the data interval radiation intensity averaging suggested by Brunnett, to improve statistical accuracy of data

10/556,918 Art Unit: 2882

collection, as suggested by Brunnett (Column 6, Lines 17-28, and Column 7, Lines 55-57) for improved imaging.

With respect to Claim 2, Yu further teaches that signal producing means includes a digital counter circuit (70) for detecting a start of a first measured data interval and a start of a next data interval (Column 6, Lines 26-40).

With respect to Claim 3, Yu further teaches that determining means further includes:

- a means (70, 72) for storing a first digital data pulse count in a first start data location and storing a first time signal value (74) associated with the first digital data pulse count in a first start time location (70) each time a pulse occurs on the digital data signal until the first measured data interval starts (Column 6, Lines 26-40), and for storing a second digital data pulse count in an end data location and storing a second time signal value (80) associated with the second digital data pulse count in an end time location (72) when the next pulse occurs on the digital data signal after the start of the next data interval is detected (Column 6, Lines 26-46);
- wherein the determining means (70, 72) determines the intensity of the detected radiation for the first measured data interval (Abstract, Lines 1-6).

Further with respect to Claim 3, Brunnett further teaches determining average intensity is achieved by dividing a difference between the pulse counts stored in the

10/556,918 Art Unit: 2882

start and end data locations by a difference between the values stored in the start and end time locations (Column 6, Lines 42-68, and Column 7, Lines 1-22).

With respect to Claims 4 and 5, Yu further teaches the converting means further includes:

- a means (44) for adding a minimized offset signal to the analog data signal so that the intensity of the analog signal is such that at least one aperiodic pulse occurs on the digital data signal during each data interval (Column 6, Lines 1-8);
- wherein the first and second data intervals are adjacent to each other (Column 6, Lines 26-46).

Allowable Subject Matter

Claims 7-20 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

With respect to Claim 7, the prior art of record teaches many of the elements of the claimed invention, including a method of measuring intensity of detected radiation in a CT scanner, the method comprising: rotating a radiation source around an examination region; generating an analog data signal that varies with an intensity of radiation traversing the region; converting the analog data signal to a digital data signal including aperiodic pulses varying in frequency with the intensity of the radiation traversing the examination region as the radiation source rotates about the examination

10/556,918 Art Unit: 2882

region; producing a time signal indicative of data intervals; determining average radiation intensity in each data interval by counting the pulses of the digital data signal starting with a digital data signal pulse occurring in a preceding data interval and continuing to a digital data signal pulse occurring in a succeeding data interval, and storing the average radiation intensity; storing a first digital data pulse count in a first start data location and storing a first time signal value in a first start time location each time a pulse occurs on the digital data signal until a first measured data interval starts; detecting a start of the first measured data interval and detecting a start of a next data interval; and determining an average intensity of the detected radiation for the first measured data interval by dividing a difference between the pulse count stored in a first data end data location and the pulse count stored in the first data start location by a difference between the value stored in the end data location and the value stored in the first start time location.

However, prior art fails to teach or fairly suggest the method wherein after said detection of next data interval, storing a second digital data signal pulse count in the end data location and storing a second time signal value in an end time location when the pulse occurs on the digital data signal, and using this second data end time location for determining the average intensity, in the manner required by Claim 7.

With respect to Claim 10, the prior art of record teaches many of the elements of the claimed invention, including a method of measuring intensity of detected radiation in a CT scanner, the method comprising: rotating a radiation source around an examination region; generating an analog data signal that varies with an intensity of

10/556,918 Art Unit: 2882

radiation traversing the region; converting the analog data signal to a digital data signal including aperiodic pulses varying in frequency with the intensity of the radiation traversing the examination region as the radiation source rotates about the examination region; producing a time signal indicative of data intervals; adding a minimized offset signal to the analog data signal prior to the converting so that the intensity of the analog data signal is such that at least one aperiodic pulse occurs on the digital data signal; and determining average radiation intensity in each data interval by counting the pulses of the digital data signal starting with a digital data signal pulse occurring in a preceding data interval and continuing to a digital data signal pulse occurring in a succeeding data interval, and storing the average radiation intensity.

However, prior art fails to teach or fairly suggest the method wherein the offset signal insures that the aperiodic pulses occurring on the digital data signal occur every 2-1/2 data intervals, in the manner required by Claim 10.

Claims 8, 9, and 11-20 would be allowable by virtue of their dependency upon Claims 7 and 10.

Response to Arguments

Applicant's arguments with respect to claims 1-6 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

U.S. Patent documents to: Kawakami (US 5,155,752), and Brunnett et al. (US 4,008,400) teach apparatuses and methods for improving image quality in an x-ray scanner using data output modulation techniques and/or radiation intensity averaging over data intervals.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Anastasia Midkiff whose telephone number is 571-272-5053. The examiner can normally be reached on M-F 7-4.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Edward Glick can be reached on 571-272-2490. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

10/556,918 Art Unit: 2882 Page 12

If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

ASM 12/20/07

EDWARD STOK SUPERVISORY PATENT CAMINER